



RIVERA MARKETING

206, Arihant Complex, Swastik Society, B/h. St. Xavier's Ladies Hostel,
Navarangpura, Ahmedabad - 380 009 ☎ 079 - 26464337
■ sales@riveramarketing.in ■ www.riveramarketing.in

Complaint no. 1669

SERVICE REPORT

Report No. 054

Date 22/06/2017

Customer's Name & Address UNITED TECHSOL (PVT) LTD HEAR MIZOLI VILAGE DHEKAD ROAD AHMEDABAD	Ser. Engr's Name : Jayesh Chaudhary
Contact Person Mr. Dheeraj Shrivastava	Machine Model : ZKD-2006
Contact Number	Serial No. : 7026-06/16
	Period of Service :
	Hours Run :

Nature of Service : ☐ Commissioning ☐ Inspection ☐ Repairs

☐ Product Complaint / Failure

Work Done : The RUC-2006 Gas Leakage
I have entered the side check valve machine
The RUC-2006 Gas is Leakage and sensor is fault so I have
checked the leakage and feel up to hydrogen pressure
and checked the leakage and find other leakage point
Dheeraj I have Rectified the RUC-2006 Gas Leakage
the sensor
Tested the unit Found working good.

Spares Used sensor, hydrogen cylinder L.P.G. vacuum pump RUC-2006	Spares Recommended
---	---------------------------

Customer's Remarks : Your Mr. _____ Called on _____ He attended to
the Job from _____ to _____ The Works in completed/not completed satisfactorily.

Service Engr. Jayesh Chaudhary 9979847531 Signature	Customer Signature & Seal of the firm
--	---

Authorised Dealers For :



GEM Equipments (P) Limited



Fine Filters Pvt Ltd.